

ABSTRACT

A method for calibrating is provided, including receiving, for each of a plurality of frequencies, an indication of a characteristic sensitivity of a position sensor for placement in a patient; measuring an actual sensitivity of the position
5 sensor at each of the plurality of frequencies; and determining, at each of the plurality of frequencies, calibration data indicative of a deviation of the actual sensitivity from the characteristic sensitivity.